

UKCA Declaration of Conformity

Mahr GmbH
Carl-Mahr-Str. 1
D - 37073 Göttingen
Germany

Mahr UK Ltd
19 Drakes Mews, Crownhill, Milton Keynes
MK8 0ER
United Kingdom

We declare under our sole responsibility that the product

Name	Height Gage
Type	Digimar 817CLT (350, 600, 1000 mm), 816CLT (350, 600 mm)
Item number	4429600, 4429601, 4429602, 4429630, 4429631
From serial number	42010001
Is in conformity with the following standards	EN ISO 12100:2010 EN ISO 55011:2016+A11:2020 EN 60204-1:2018 EN IEC 61000-6-1:2019 EN IEC 61000-6-2:2019 EN IEC 61000-6-3:2021 EN IEC 61000-6-4:2019 EN 61010-1:2010/A1:2019 EN IEC 63000:2018
Following the Directive(s)	2008 No 1597 2016 No 1091 2012 No 422

Technical documentation team leader


Stefan Günther, Mahr GmbH

Executive director

Manuel Hüsken

Place and Date

Göttingen, 2023-12-15



.....
Manuel Hüsken